

Inflammation Markers Control – Level 1

Assignment of values (**LOT** 07C0901)

Instrument	Analyte	Unit	Inflammation Markers Control – Level 1		
			Target	SD	range
AutoLumo A2000 Plus	PCT	ng/mL	0.50	0.07	0.30-0.71
	IL-6	pg/mL	15.47	1.73	10.28-20.66
	CRP	mg/L	5.55	0.51	4.02-7.08
Roche cobas 6000 e 601	PCT	ng/mL	0.25	0.02	0.19-0.31
	IL-6	pg/mL	20.62	2.61	12.77-28.46
Beckman Coulter AU680	CRP	mg/L	5.76	0.82	3.31-8.21
	SAA	mg/L	9.98	1.06	6.79-13.16
	MPO	ng/mL	57.47	7.26	35.68-79.25
	Lp-PLA2	U/L	325.05	46.38	185.90-464.20
Canon TBA-FX8/TBA-120FR	CRP	mg/L	5.31	0.74	3.08-7.54
	SAA	mg/L	10.70	1.22	7.04-14.36
	MPO	ng/mL	50.68	6.60	30.88-70.47
	Lp-PLA2	U/L	331.13	49.03	184.03-478.22

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.