

Inflammation Markers Control – Level 2

Assignment of values (**LOT** 07C0901)

Instrument	Analyte	Unit	Inflammation Markers Control – Level 2		
			Target	SD	range
AutoLumo A2000 Plus	PCT	ng/mL	2.11	0.22	1.44-2.79
	IL-6	pg/mL	71.63	8.46	46.25-97.01
	CRP	mg/L	16.80	1.71	11.67-21.93
Roche cobas 6000 e 601	PCT	ng/mL	1.21	0.12	0.85-1.56
	IL-6	pg/mL	85.05	8.31	60.13-109.97
Beckman Coulter AU680	CRP	mg/L	16.25	1.34	12.23-20.27
	SAA	mg/L	18.92	2.00	12.91-24.93
	MPO	ng/mL	125.45	12.79	87.07-163.83
	Lp-PLA2	U/L	478.75	71.26	264.98-692.52
Canon TBA-FX8/TBA-120FR	CRP	mg/L	16.22	1.96	10.35-22.09
	SAA	mg/L	19.61	2.52	12.04-27.17
	MPO	ng/mL	118.04	14.56	74.35-161.72
	Lp-PLA2	U/L	488.42	55.32	322.45-654.38

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.