

Inflammation Markers Control – Level 3

Assignment of values (**LOT** 07C0901)

Instrument	Analyte	Unit	Inflammation Markers Control – Level 3		
			Target	SD	range
AutoLumo A2000 Plus	PCT	ng/mL	23.78	3.31	13.85-33.70
	IL-6	pg/mL	450.51	48.97	303.60-597.42
	CRP	mg/L	34.91	3.37	24.80-45.03
Roche cobas 6000 e 601	PCT	ng/mL	16.57	1.45	12.24-20.91
	IL-6	pg/mL	535.03	52.29	378.16-691.89
Beckman Coulter AU680	CRP	mg/L	32.05	3.47	21.64-42.46
	SAA	mg/L	36.76	4.17	24.25-49.28
	MPO	ng/mL	708.75	76.56	479.06-938.43
	Lp-PLA2	U/L	895.30	110.43	564.01-1227
Canon TBA-FX8/TBA-120FR	CRP	mg/L	31.99	3.89	20.33-43.64
	SAA	mg/L	37.52	4.74	23.30-51.73
	MPO	ng/mL	659.27	84.07	407.05-911.48
	Lp-PLA2	U/L	919.80	106.01	601.77-1238

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.