

# Infectious Disease Control III– Level 1

Assignment of values ( **REF** BYZKG0110 **LOT** 04L1003 )

Instrument	Analyte	Unit	Infectious Disease Control
			III-Level 1 Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900	HBsAg	IU/mL	<0.05
	HBeAg	IU/mL	<0.10
	Anti-TP	S/CO	<1.00
	Anti-HCV	S/CO	<1.00
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	<1.00
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	<1.00
Auto Lumo A1860	HBsAg	IU/mL	<0.05
	HBeAg	IU/mL	<0.10
	Anti-TP	S/CO	<1.00
	Anti-HCV	S/CO	<1.00
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	<1.00
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	<1.00

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.