

# Infectious Disease Control IV– Level 3

Assignment of values ( **REF** BYZKG1310 **LOT** 18C1101 )

Instrument	Analyte	Unit	Infectious Disease Control IV-Level 3	
			Target	Range
Auto Lumo A2000/Auto Lumo A2000 Plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900	Anti-HBs	mIU/mL	111.70	67.02-156.38
	Anti-HBe	PEI U/mL	1.03	0.62-1.44
	Anti-HBc	PEI U/mL	5.08	3.05-7.12
	HIV p24 Ag	S/CO	5.83	3.50-8.17
Auto Lumo A1860	Anti-HBs	mIU/mL	114.47	68.68-160.26
	Anti-HBe	PEI U/mL	1.09	0.65-1.52
	Anti-HBc	PEI U/mL	5.27	3.16-7.38
	HIV p24 Ag	S/CO	5.65	3.39-7.92

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.