

Infectious Disease Control IV– Level 1

Assignment of values (**REF** BYZKG1110 **LOT** 18C1101)

Instrument	Analyte	Unit	Infectious Disease Control IV-Level 1
			Range
Auto Lumo A2000/Auto Lumo A2000 Plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900	Anti-HBs	mIU/mL	<10.00
	Anti-HBe	PEI U/mL	<0.15
	Anti-HBc	PEI U/mL	<0.70
	HIV p24 Ag	S/CO	<1.00
Auto Lumo A1860	Anti-HBs	mIU/mL	<10.00
	Anti-HBe	PEI U/mL	<0.15
	Anti-HBc	PEI U/mL	<0.70
	HIV p24 Ag	S/CO	<1.00

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.