

Infectious Disease Control III– Level 3

Assignment of values (**REF** BYZKG0310 **LOT** 24C1001)

Instrument	Analyte	Unit	Infectious Disease Control III-Level 3	
			Target	Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900	HBsAg	IU/mL	3.68	2.21-5.15
	HBeAg	IU/mL	3.09	1.86-4.33
	Anti-TP	S/CO	5.70	3.42-7.98
	Anti-HCV	S/CO	4.42	2.65-6.19
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	6.58	3.95-9.21
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	5.97	3.58-8.36
Auto Lumo A1860	HBsAg	IU/mL	3.57	2.14-5.00
	HBeAg	IU/mL	2.90	1.74-4.06
	Anti-TP	S/CO	5.64	3.39-7.90
	Anti-HCV	S/CO	4.73	2.84-6.62
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	6.14	3.68-8.60
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	5.92	3.55-8.29

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.