

# Infectious Disease Control IV– Level 2

Assignment of values ( **REF** BYZKG1210 **LOT** 23G0902 )

Instrument	Analyte	Unit	Infectious Disease Control IV-Level 2	
			Target	Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900	Anti-HBs	mIU/mL	34.90	20.94-48.86
	Anti-HBe	PEI U/mL	0.50	0.30-0.70
	Anti-HBc	PEI U/mL	2.17	1.30-3.04
	HIV p24 Ag	S/CO	2.76	1.66-3.87
Auto Lumo A1860	Anti-HBs	mIU/mL	34.83	20.90-48.76
	Anti-HBe	PEI U/mL	0.47	0.28-0.66
	Anti-HBc	PEI U/mL	2.01	1.21-2.81
	HIV p24 Ag	S/CO	3.03	1.82-4.24

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.