

Infectious Disease Control IV– Level 4

Assignment of values (**REF** BYZKG1410 **LOT** 23G0901)

Instrument	Analyte	Unit	Infectious Disease Control IV-Level 4	
			Target	Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900	Anti-HBs	mIU/mL	230.43	138.26-322.61
	Anti-HBe	PEI U/mL	2.47	1.48-3.45
	Anti-HBc	PEI U/mL	9.80	5.88-13.71
	HIV p24 Ag	S/CO	11.54	6.93-16.16
Auto Lumo A1860	Anti-HBs	mIU/mL	252.59	151.56-353.63
	Anti-HBe	PEI U/mL	2.29	1.37-3.21
	Anti-HBc	PEI U/mL	9.62	5.77-13.47
	HIV p24 Ag	S/CO	12.90	7.74-18.06

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.