

## Inflammation Markers Control – Level 3

Assignment of values ( **REF** BYZKB0301 **LOT** 07B1101 )

Instrument	Analyte	Unit	Inflammation Markers Control – Level 3		
			Target	SD	range
AutoLumo A2000 Plus	PCT	ng/mL	24.95	2.80	16.56-33.34
	IL-6	pg/mL	486.04	46.50	346.53-625.55
	CRP	mg/L	34.68	4.02	22.61-46.74
Roche cobas 6000 e 601	PCT	ng/mL	17.27	2.04	11.15-23.38
	IL-6	pg/mL	571.25	53.00	412.24-730.26
Beckman Coulter AU680	CRP	mg/L	31.96	3.34	21.93-41.99
	Lp-PLA2	U/L	818.73	79.84	579.22-1058
	MPO	ng/mL	735.56	69.99	525.59-945.52
	SAA	mg/L	34.17	3.77	22.86-45.47
Canon TBA-FX8/TBA-120FR	CRP	mg/L	34.16	3.96	22.28-46.04
	Lp-PLA2	U/L	827.44	74.87	602.84-1052
	MPO	ng/mL	674.75	67.51	472.22-877.28
	SAA	mg/L	32.27	3.69	21.21-43.33

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.