

Inflammation Markers Control – Level 1

Assignment of values (**REF** BYZKB0101 **LOT** 13K1002)

Instrument	Analyte	Unit	Inflammation Markers Control – Level 1		
			Target	SD	range
AutoLumo A2000 Plus	PCT	ng/mL	0.50	0.05	0.35-0.64
	IL-6	pg/mL	16.77	1.74	11.57-21.98
	CRP	mg/L	5.52	0.63	3.63-7.41
Roche cobas 6000 e 601	PCT	ng/mL	0.24	0.02	0.17-0.30
	IL-6	pg/mL	21.68	2.72	13.53-29.83
Beckman Coulter AU680	CRP	mg/L	6.59	0.83	4.11-9.07
	Lp-PLA2	U/L	290.68	30.24	199.96-381.40
	MPO	ng/mL	53.70	6.91	32.97-74.43
	SAA	mg/L	10.39	1.10	7.10-13.68
Canon TBA-FX8/TBA-120FR	CRP	mg/L	5.16	0.53	3.57-6.75
	Lp-PLA2	U/L	296.76	32.37	199.64-393.88
	MPO	ng/mL	53.60	5.68	36.57-70.63
	SAA	mg/L	10.03	1.18	6.48-13.57

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.