

## AMH Control – Level 1

Assignment of values ( **REF** BYZKA0102 **LOT** 25H1002 )

Instrument	Analyte	Unit	AMH Control-Level1		
			Target	SD	range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	0.90	0.14	0.63-1.17
Roche Cobas e601	AMH	pmol/L	6.24	0.94	4.37-8.11
Beckman DXI800	AMH	ng/mL	0.86	0.13	0.60-1.12

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.