

Inflammation Markers Control – Level 2

Assignment of values (**REF** BYZKB0201 **LOT** 13K1002)

Instrument	Analyte	Unit	Inflammation Markers Control – Level 2		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	2.21	0.21	1.59-2.83
	IL-6	pg/mL	68.05	6.57	48.34-87.76
	CRP	mg/L	16.68	1.93	10.90-22.46
Roche cobas 6000 e 601	PCT	ng/mL	1.21	0.11	0.87-1.56
	IL-6	pg/mL	84.97	8.06	60.78-109.16
Beckman Coulter AU680	CRP	mg/L	17.16	1.87	11.54-22.77
	Lp-PLA2	U/L	403.68	48.49	258.22-549.15
	MPO	ng/mL	131.33	14.73	87.15-175.51
	SAA	mg/L	18.03	1.74	12.80-23.26
Canon TBA-FX8/TBA-120FR	CRP	mg/L	15.98	1.61	11.14-20.82
	Lp-PLA2	U/L	414.94	39.25	297.19-532.68
	MPO	ng/mL	128.93	13.14	89.51-168.35
	SAA	mg/L	17.80	1.85	12.25-23.34

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.