

Inflammation Markers Control – Level 3

Assignment of values (**REF** BYZKB0301 **LOT** 13K1002)

Instrument	Analyte	Unit	Inflammation Markers Control – Level 3		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	26.44	3.07	17.24-35.64
	IL-6	pg/mL	465.32	43.49	334.86-595.78
	CRP	mg/L	33.45	3.79	22.09-44.80
Roche cobas 6000 e 601	PCT	ng/mL	17.38	1.85	11.83-22.94
	IL-6	pg/mL	549.89	54.36	386.80-712.98
Beckman Coulter AU680	CRP	mg/L	32.40	3.24	22.69-42.10
	Lp-PLA2	U/L	802.80	75.41	576.56-1029
	MPO	ng/mL	730.00	88.78	463.65-996.35
	SAA	mg/L	35.46	4.24	22.75-48.18
Canon TBA-FX8/TBA-120FR	CRP	mg/L	32.23	3.21	22.59-41.87
	Lp-PLA2	U/L	817.63	81.80	572.23-1063
	MPO	ng/mL	678.50	68.80	472.10-884.90
	SAA	mg/L	35.52	3.49	25.06-45.98

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.