

Inflammation Markers Control – Level 1

Assignment of values (**REF** BYZKB0101 **LOT** 07B1101)

Instrument	Analyte	Unit	Inflammation Markers Control – Level 1		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	0.48	0.04	0.35-0.61
	IL-6	pg/mL	15.33	1.84	9.80-20.87
	CRP	mg/L	5.42	0.53	3.84-7.00
Roche cobas 6000 e 601	PCT	ng/mL	0.23	0.03	0.15-0.31
	IL-6	pg/mL	19.40	2.29	12.54-26.26
Beckman Coulter AU680	CRP	mg/L	5.78	0.73	3.59-7.97
	Lp-PLA2	U/L	292.99	33.79	191.63-394.36
	MPO	ng/mL	58.78	6.56	39.10-78.46
	SAA	mg/L	10.55	1.03	7.48-13.63
Canon TBA-FX8/TBA-120FR	CRP	mg/L	4.76	0.53	3.16-6.35
	Lp-PLA2	U/L	291.90	33.06	192.72-391.07
	MPO	ng/mL	54.98	5.58	38.25-71.70
	SAA	mg/L	10.19	1.09	6.92-13.45

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.