

## AMH Control – Level 3

Assignment of values ( **REF** BYZKA0302 **LOT** 04M0803 )

Instrument	Analyte	Unit	AMH Control Level3		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	9.64	1.45	6.75-12.53
Roche Cobas e601	AMH	pmol/L	69.24	10.39	48.47-90.01
Beckman DXI800	AMH	ng/mL	9.45	1.42	6.62-12.29

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.