

AMH Control – Level 2

Assignment of values (**REF** BYZKA0202 **LOT** 04M0803)

Instrument	Analyte	Unit	AMH Control Level2		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	4.48	0.67	3.14-5.82
Roche Cobas e601	AMH	pmol/L	31.93	4.79	22.35-41.51
Beckman DXI800	AMH	ng/mL	4.15	0.62	2.91-5.40

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.