

AMH Control – Level 1

Assignment of values (**REF** BYZKA0102 **LOT** 04M0803)

Instrument	Analyte	Unit	AMH Control Level1		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	0.92	0.14	0.64-1.20
Roche Cobas e601	AMH	pmol/L	6.39	0.96	4.47-8.30
Beckman DXI800	AMH	ng/mL	0.86	0.13	0.60-1.12

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.