

## AMH Control – Level 2

Assignment of values ( **REF** BYZKA0202 **LOT** 20H0802 )

Instrument	Analyte	Unit	AMH Control Level2		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	5.02	0.75	3.51-6.52
Roche Cobas e601	AMH	pmol/L	29.57	4.44	20.70-38.44
Beckman DXI800	AMH	ng/mL	3.84	0.58	2.69-4.99
iFlash 3000-C	AMH	ng/mL	8.29	1.24	5.80-10.78

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.