

AMH Control – Level 3

Assignment of values (**REF** BYZKA0302 **LOT** 20H0802)

Instrument	Analyte	Unit	AMH Control Level3		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	9.71	1.46	6.80-12.63
Roche Cobas e601	AMH	pmol/L	61.08	9.16	42.75-79.40
Beckman DXI800	AMH	ng/mL	8.34	1.25	5.84-10.84
iFlash 3000-C	AMH	ng/mL	15.20	2.28	10.64-19.76

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.