

AMH Control – Level 1

Assignment of values (**REF** BYZKA0102 **LOT** 20H0802)

Instrument	Analyte	Unit	AMH Control Level1		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	1.04	0.16	0.73-1.35
Roche Cobas e601	AMH	pmol/L	4.88	0.73	3.42-6.35
Beckman DXI800	AMH	ng/mL	0.66	0.10	0.46-0.86
iFlash 3000-C	AMH	ng/mL	1.68	0.25	1.18-2.18

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.