

AMH Control – Level 3

Assignment of values (**REF** BYZKA0302 **LOT** 22G0901)

Instrument	Analyte	Unit	AMH Control Level3		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	9.27	1.39	6.49-12.05
Roche Cobas e601	AMH	pmol/L	75.57	11.34	52.90-98.24
Beckman DXI800	AMH	ng/mL	8.37	1.26	5.86-10.88

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.