

AMH Control – Level 1

Assignment of values (**REF** BYZKA0102 **LOT** 22G0901)

| Instrument | Analyte | Unit | AMH Control Level1 | | |
|----------------------------|---------|--------|--------------------|------|-----------|
| | | | Target | SD | Range |
| Auto Lumo A2000/A2000 plus | AMH | ng/mL | 0.82 | 0.12 | 0.57-1.07 |
| Roche Cobas e601 | AMH | pmol/L | 6.59 | 0.99 | 4.61-8.57 |
| Beckman DXI800 | AMH | ng/mL | 0.68 | 0.10 | 0.48-0.88 |

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.