

AMH Control – Level 2

Assignment of values (**REF** BYZKA0202 **LOT** 22G0901)

Instrument	Analyte	Unit	AMH Control Level2		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	4.26	0.64	2.98-5.54
Roche Cobas e601	AMH	pmol/L	34.57	5.19	24.20-44.94
Beckman DXI800	AMH	ng/mL	4.07	0.61	2.85-5.29

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.