

AMH Control – Level 1

Assignment of values (**REF** BYZKA0102 **LOT** 25H1001)

| Instrument | Analyte | Unit | AMH Control Level1 | | |
|----------------------------|---------|--------|--------------------|------|-----------|
| | | | Target | SD | Range |
| Auto Lumo A2000/A2000 plus | AMH | ng/mL | 0.90 | 0.14 | 0.63-1.17 |
| Roche Cobas e601 | AMH | pmol/L | 6.24 | 0.94 | 4.37-8.11 |
| Beckman DXI800 | AMH | ng/mL | 0.86 | 0.13 | 0.60-1.12 |

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.