

## AMH Control – Level 2

Assignment of values ( **REF** BYZKA0202 **LOT** 25H1002 )

Instrument	Analyte	Unit	AMH Control Level2		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	4.49	0.68	3.14-5.84
Roche Cobas e601	AMH	pmol/L	32.35	4.85	22.65-42.06
Beckman DXI800	AMH	ng/mL	4.15	0.62	2.91-5.40

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.