

## Specific Protein Control– Level 2

Assignment of values ( **REF** CAB0202    **LOT** 18D1001 )

Instrument	Analyte	Unit	Specific Protein Control-Level2		
			Target	SD	Range
Toshiba TBA-120FR /TBA-FX8	ALB	g/L	51.04	4.85	36.50-65.58
	TP	g/L	83.96	8.68	57.91-110.01
	IgA	mg/dL	273.54	27.13	192.15-354.92
	IgG	mg/dL	1261	117.08	910-1612
	IgM	mg/dL	128.05	12.45	90.69-165.41
	IgE	IU/mL	227.74	22.27	160.92-294.55
	C3	mg/dL	169.56	16.20	120.95-218.16
	C4	mg/dL	33.34	3.52	22.79-43.90
	RF	U/mL	40.64	3.64	29.72-51.57
	ASO	U/mL	199.02	22.90	130.31-267.73
	CRP	mg/L	31.87	3.50	21.36-42.37
	Ferr	ng/mL	254.84	28.54	169.24-340.45
	TRF	g/L	3.12	0.34	2.09-4.16
	RBP	mg/L	62.79	6.86	42.21-83.38
	α1-MG	mg/L	32.19	2.91	23.47-40.91
	β2-MG	mg/L	2.17	0.25	1.41-2.94
	PAB	mg/L	290.87	31.09	197.60-384.14
	CysC	mg/L	2.05	0.23	1.37-2.73
	HPT	g/L	1.34	0.14	0.93-1.74
	AAT	g/L	1.58	0.18	1.03-2.12
κ-LC	g/L	3.25	0.31	2.34-4.17	
λ-LC	g/L	2.03	0.20	1.44-2.62	
Beckman AU680	ALB	g/L	51.47	4.23	38.78-64.17
	TP	g/L	77.53	7.00	56.53-98.53
	IgA	g/L	2.85	0.24	2.12-3.58
	IgG	g/L	13.42	1.34	9.39-17.45
	IgM	g/L	1.10	0.12	0.73-1.47
	IgE	IU/mL	212.24	21.43	147.96-276.52
	C3	g/L	1.75	0.19	1.18-2.31
	C4	g/L	0.33	0.04	0.22-0.44
	RF	IU/mL	38.54	3.03	29.44-47.64
	ASO	IU/mL	203.69	16.55	154.04-253.35
	CRP	mg/L	33.32	2.61	25.49-41.15
	Ferr	ng/mL	224.44	24.16	151.95-296.92
	TRF	g/L	2.92	0.25	2.16-3.68

Instrument	Analyte	Unit	Specific Protein Control-Level2		
			Target	SD	Range
Beckman AU680	RBP	mg/L	46.57	3.31	36.65-56.50
	$\alpha$ 1-MG	mg/L	26.78	2.31	19.84-33.72
	$\beta$ 2-MG	mg/L	2.03	0.19	1.45-2.62
	PAB	mg/L	291.83	27.48	209.38-374.27
	CysC	mg/L	1.61	0.16	1.12-2.10
	HPT	g/L	1.30	0.10	1.01-1.58

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.