

AMH Control – Level 3

Assignment of values (**REF** BYZKA0302 **LOT** 25H1003)

Instrument	Analyte	Unit	AMH Control Level3		
			Target	SD	Range
Auto Lumo A2000/A2000 plus	AMH	ng/mL	8.91	1.34	6.24-11.58
Roche Cobas e601	AMH	pmol/L	64.30	9.65	45.01-83.59
Beckman DXI800	AMH	ng/mL	7.99	1.20	5.59-10.39

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.