

Coagulation Control – Level 3

Assignment of values (**REF** BYZKF0303 **LOT** 16C1101)

Instrument	Analyte	Unit	Coagulation Control - Level3		
			Target	SD	Range
Sysmex CS 2000i	PT	s	28.16	1.69	23.08-33.23
	APTT	s	70.48	3.18	60.94-80.02
	TT	s	28.65	1.36	24.58-32.72
	FIB	mg/dL	109.79	7.77	86.47-133.11
	AT III	%	40.37	2.00	34.37-46.37
	DD	mg/L	4.14	0.30	3.24-5.04
	FDP	mg/L	13.03	1.10	9.74-16.31
Syscan Ci 300	PT	s	24.52	1.25	20.77-28.27
	APTT	s	73.94	2.71	65.80-82.07
	TT	s	28.15	1.34	24.13-32.17
	FIB	mg/dL	139.05	5.86	121.46-156.64
	AT III	%	40.93	2.07	34.73-47.13
	DD	mg/L	3.89	0.16	3.42-4.36
	FDP	mg/L	13.91	0.90	11.22-16.60
IL ACL TOP 750	PT	s	28.00	1.46	23.61-32.39
	APTT	s	59.78	2.21	53.16-66.39
	TT	s	31.53	2.35	24.47-38.59
	FIB	mg/dL	131.98	8.91	105.24-158.71
	AT III	%	49.24	2.05	43.11-55.38
	DD	mg/L	0.84	0.10	0.53-1.15
	FDP	mg/L	5.78	0.31	4.86-6.70
积水 CP3000	PT	s	31.87	2.13	25.49-38.24
	APTT	s	75.57	5.04	60.45-90.68
	TT	s	27.73	1.85	22.19-33.28
	FIB	mg/dL	183.33	12.22	146.67-220.00
	AT III	%	40.47	2.70	32.37-48.56
	DD	mg/L	4.84	0.32	3.87-5.81
	FDP	mg/L	8.80	0.59	7.04-10.56

Instrument	Analyte	Unit	Coagulation Control - Level3		
			Target	SD	Range
迈瑞 CX-9010	PT	s	26.08	2.61	18.26-33.90
	APTT	s	76.60	7.66	53.62-99.58
	TT	s	28.26	2.83	19.78-36.73
	FIB	mg/dL	1.57	0.16	1.10-2.04
	DD	mg/L	3.77	0.38	2.64-4.90

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.