

Coagulation Control – Level 1

Assignment of values (**REF** BYZKF0103 **LOT** 16C1101)

Instrument	Analyte	Unit	Coagulation Control - Level1		
			Target	SD	Range
Sysmex CS 2000i	PT	s	10.92	0.57	9.20-12.63
	APTT	s	26.80	1.33	22.80-30.80
	TT	s	22.01	1.07	18.80-25.21
	FIB	mg/dL	311.33	21.81	245.91-376.74
	AT III	%	75.71	2.72	67.54-83.87
	DD	mg/L	0.41	0.04	0.30-0.52
	FDP	mg/L	3.63	0.46	2.26-4.99
Syscan Ci 300	PT	s	9.59	0.54	7.98-11.19
	APTT	s	29.92	1.45	25.58-34.25
	TT	s	19.18	0.90	16.47-21.88
	FIB	mg/dL	246.50	11.56	211.81-281.19
	AT III	%	77.34	3.60	66.53-88.14
	DD	mg/L	0.61	0.08	0.37-0.86
	FDP	mg/L	3.39	0.60	1.59-5.18
IL ACL TOP 750	PT	s	10.55	0.64	8.62-12.47
	APTT	s	29.70	1.59	24.94-34.46
	TT	s	14.85	0.89	12.19-17.50
	FIB	mg/dL	283.99	12.42	246.74-321.25
	AT III	%	83.46	3.67	72.45-94.46
	DD	mg/L	0.18	0.01	0.15-0.22
	FDP	mg/L	0.64	0.20	0.03-1.25
积水 CP3000	PT	s	9.87	0.66	7.89-11.84
	APTT	s	30.70	2.05	24.56-36.84
	TT	s	16.97	1.13	13.57-20.36
	FIB	mg/dL	310.00	20.67	248.00-372.00
	AT III	%	69.43	4.63	55.55-83.32
	DD	mg/L	0.76	0.05	0.61-0.91
	FDP	mg/L	2.53	0.17	2.03-3.04

Instrument	Analyte	Unit	Coagulation Control - Level1		
			Target	SD	Range
迈瑞 CX-9010	PT	s	10.06	1.01	7.04-13.08
	APTT	s	32.63	3.26	22.84-42.42
	TT	s	14.41	1.44	10.09-18.73
	FIB	mg/dL	2.55	0.26	1.79-3.32
	DD	mg/L	0.52	0.05	0.36-0.68

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.