

Coagulation Control – Level 2

Assignment of values (**REF** BYZKF0203 **LOT** 16C1101)

Instrument	Analyte	Unit	Coagulation Control - Level2		
			Target	SD	Range
Sysmex CS 2000i	PT	s	18.57	0.82	16.12-21.01
	APTT	s	48.75	2.11	42.43-55.07
	TT	s	25.03	1.26	21.25-28.80
	FIB	mg/dL	147.25	6.04	129.12-165.38
	AT III	%	56.14	2.26	49.36-62.92
	DD	mg/L	2.38	0.10	2.08-2.67
	FDP	mg/L	9.35	0.78	7.02-11.68
Syscan Ci 300	PT	s	16.78	1.04	13.67-19.89
	APTT	s	46.43	2.10	40.14-52.71
	TT	s	22.57	1.20	18.98-26.15
	FIB	mg/dL	177.00	8.72	150.85-203.15
	AT III	%	57.66	2.31	50.74-64.57
	DD	mg/L	2.51	0.20	1.91-3.10
	FDP	mg/L	9.48	0.53	7.89-11.07
IL ACL TOP 750	PT	s	16.42	0.76	14.15-18.69
	APTT	s	35.05	1.99	29.08-41.02
	TT	s	23.74	1.24	20.01-27.46
	FIB	mg/dL	161.67	9.01	134.65-188.69
	AT III	%	67.28	2.57	59.58-74.98
	DD	mg/L	0.54	0.06	0.37-0.70
	FDP	mg/L	3.41	0.27	2.59-4.23
积水 CP3000	PT	s	16.63	1.11	13.31-19.96
	APTT	s	52.50	3.50	42.00-63.00
	TT	s	22.20	1.48	17.76-26.64
	FIB	mg/dL	263.67	17.58	210.93-316.40
	AT III	%	53.50	3.57	42.80-64.20
	DD	mg/L	2.70	0.18	2.16-3.24
	FDP	mg/L	5.60	0.37	4.48-6.72

Instrument	Analyte	Unit	Coagulation Control - Level2		
			Target	SD	Range
迈瑞 CX-9010	PT	s	14.55	1.46	10.19-18.92
	APTT	s	48.82	4.88	34.17-63.46
	TT	s	23.02	2.30	16.11-29.92
	FIB	mg/dL	2.18	0.22	1.53-2.84
	DD	mg/L	2.14	0.21	1.50-2.78

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.