

Infectious Disease Control IV – Level 4

Assignment of values (**REF** BYZKG1410 **LOT** 21E1103)

Instrument	Analyte	Unit	Infectious Disease Control IV-Level 4		
			Target	SD	Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200	Anti-HBs	mIU/mL	213.77	28.50	128.26-299.28
	Anti-HBe	PEI U/mL	2.00	0.27	1.20-2.80
	Anti-HBc	PEI U/mL	10.13	1.35	6.08-14.18
	HIV p24 Ag	S/CO	12.37	1.65	7.42-17.32

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.