

Infectious Disease Control IV – Level 2

Assignment of values (**REF** BYZKG1210 **LOT** 20E1102)

Instrument	Analyte	Unit	Infectious Disease Control IV-Level 2		
			Target	SD	Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200	Anti-HBs	mIU/mL	30.97	4.13	18.58-43.36
	Anti-HBe	PEI U/mL	0.45	0.06	0.27-0.63
	Anti-HBc	PEI U/mL	2.09	0.28	1.26-2.93
	HIV p24 Ag	S/CO	2.62	0.35	1.57-3.67

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.