

Infectious Disease Control III – Level 4

Assignment of values (**REF** BYZKG0410 **LOT** 21E1103)

Instrument	Analyte	Unit	Infectious Disease Control III-Level 4		
			Target	SD	Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900/Auto Lumo A1860	HBsAg	IU/mL	18.13	2.44	10.88-25.39
	HBeAg	IU/mL	6.03	0.80	3.62-8.44
	Anti-TP	S/CO	11.63	1.55	6.98-16.28
	Anti-HCV	S/CO	8.92	1.19	5.35-12.48
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	12.01	1.60	7.21-16.81
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	11.20	1.49	6.72-15.68
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles) (RU)	S/CO	5.56	0.74	3.34-7.78

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.