

Infectious Disease Control III – Level 2

Assignment of values (**REF** BYZKG0210 **LOT** 20E1102)

Instrument	Analyte	Unit	Infectious Disease Control III-Level 2		
			Target	SD	Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900/Auto Lumo A1860	HBsAg	IU/mL	0.45	0.06	0.27-0.63
	HBeAg	IU/mL	0.32	0.04	0.19-0.45
	Anti-TP	S/CO	2.92	0.39	1.75-4.08
	Anti-HCV	S/CO	2.65	0.35	1.59-3.70
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	2.95	0.39	1.77-4.13
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	2.59	0.35	1.55-3.63
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles) (RU)	S/CO	1.10	0.18	0.55-1.65

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.