

Infectious Disease Control III – Level 3

Assignment of values (**REF** BYZKG0310 **LOT** 20E1102)

Instrument	Analyte	Unit	Infectious Disease Control III-Level 3		
			Target	SD	Range
Auto Lumo A2000/Auto Lumo A2000 plus/Auto Lumo A1000/Auto Lumo A6200/AutoLumo S900/Auto Lumo A1860	HBsAg	IU/mL	3.68	0.49	2.21-5.15
	HBeAg	IU/mL	3.09	0.41	1.86-4.33
	Anti-TP	S/CO	5.70	0.76	3.42-7.98
	Anti-HCV	S/CO	4.42	0.59	2.65-6.19
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	6.58	0.88	3.95-9.21
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	5.97	0.80	3.58-8.36
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles) (RU)	S/CO	2.98	0.50	1.49-4.47

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.