

## Inflammation Markers Control – Level 1

Assignment of values ( **REF** BYZKB0101 **LOT** 10C1001 )

Instrument	Analyte	Unit	Inflammation Markers Control Level 1		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	0.43	0.05	0.30-0.57
	IL-6	pg/mL	15.17	1.63	10.28-20.07
	CRP	mg/L	5.40	0.57	3.70-7.09
Roche cobas 6000 e 601	PCT	ng/mL	0.22	0.02	0.16-0.29
	IL-6	pg/mL	20.35	1.81	14.91-25.78
Canon TBA-FX8/TBA-120FR	CRP	mg/L	4.48	0.49	3.02-5.94
	Lp-PLA2	U/L	288.28	28.78	201.95-374.60
	MPO	ng/mL	52.17	5.65	35.22-69.12
	SAA	mg/L	10.44	0.94	7.61-13.26

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.