

Inflammation Markers Control – Level 2

Assignment of values (**REF** BYZKB0201 **LOT** 10C1001)

Instrument	Analyte	Unit	Inflammation Markers Control Level 2		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	1.93	0.18	1.40-2.45
	IL-6	pg/mL	69.16	6.32	50.21-88.11
	CRP	mg/L	15.84	1.53	11.25-20.42
Roche cobas 6000 e 601	PCT	ng/mL	1.08	0.10	0.77-1.40
	IL-6	pg/mL	85.19	7.45	62.83-107.55
Canon TBA-FX8/TBA-120FR	CRP	mg/L	14.30	1.44	9.99-18.62
	Lp-PLA2	U/L	414.95	38.08	300.70-529.20
	MPO	ng/mL	118.39	11.30	84.49-152.29
	SAA	mg/L	18.65	1.84	13.13-24.16

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.