

## Inflammation Markers Control – Level 3

Assignment of values ( **REF**BYZKB0301 **LOT** 10C1001 )

Instrument	Analyte	Unit	Inflammation Markers Control Level 3		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	25.19	2.39	18.03-32.35
	IL-6	pg/mL	453.00	36.40	343.80-562.20
	CRP	mg/L	32.70	3.28	22.87-42.53
Roche cobas 6000 e 601	PCT	ng/mL	18.05	1.73	12.86-23.24
	IL-6	pg/mL	546.73	51.50	392.22-701.23
Canon TBA-FX8/TBA-120FR	CRP	mg/L	28.15	2.87	19.53-36.77
	Lp-PLA2	U/L	826.37	70.04	616.23-1036
	MPO	ng/mL	622.96	57.74	449.73-796.18
	SAA	mg/L	37.63	3.50	27.12-48.14

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.