

Inflammation Markers Control – Level 2

Assignment of values (**REF** BYZKB0201 **LOT** 07B1103)

Instrument	Analyte	Unit	Inflammation Markers Control Level2		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	2.17	0.22	1.50-2.83
	IL-6	pg/mL	62.31	6.58	42.57-82.05
	CRP	mg/L	17.02	1.74	11.80-22.23
Roche cobas 6000 e 601	PCT	ng/mL	1.24	0.12	0.90-1.59
	IL-6	pg/mL	77.62	8.88	50.97-104.27
Beckman Coulter AU680	CRP	mg/L	17.78	1.74	12.57-22.98
	Lp-PLA2	U/L	409.02	39.04	291.89-526.16
	MPO	ng/mL	143.88	17.42	91.61-196.15
	SAA	mg/L	19.26	1.95	13.40-25.13
Canon TBA-FX8/TBA-120FR	CRP	mg/L	16.89	1.61	12.07-21.72
	Lp-PLA2	U/L	407.28	41.31	283.34-531.22
	MPO	ng/mL	131.06	12.13	94.66-167.46
	SAA	mg/L	17.91	1.65	12.94-22.87

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.