

AMH Control – Level 3

Assignment of values (**REF** BYZKA0302 **LOT** 25H1004)

| Instrument | Analyte | Unit | AMH Control Level 3 | | |
|----------------------------|---------|--------|---------------------|------|-------------|
| | | | Target | SD | Range |
| Auto Lumo A2000/A2000 plus | AMH | ng/mL | 8.91 | 1.34 | 6.24-11.58 |
| Roche Cobas e601 | AMH | pmol/L | 64.30 | 9.65 | 45.01-83.59 |
| Beckman DXI800 | AMH | ng/mL | 7.99 | 1.20 | 5.59-10.39 |

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.