

Inflammation Markers Control – Level 1

Assignment of values (**REF** BYZKB0101 **LOT** 10H1104)

Instrument	Analyte	Unit	Inflammation Markers Control Level1		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	0.45	0.05	0.30-0.60
	IL-6	pg/mL	15.48	1.83	9.99-20.97
	CRP	mg/L	4.71	0.50	3.21-6.21
Roche cobas 6000 e 601	PCT	ng/mL	0.23	0.02	0.17-0.29
	IL-6	pg/mL	21.07	1.95	15.22-26.92
Beckman Coulter AU680	CRP	mg/L	4.68	0.56	3.00-6.36
	Lp-PLA2	U/L	287.99	32.45	190.64-385.34
	MPO	ng/mL	54.49	5.75	37.24-71.74
	SAA	mg/L	9.14	1.06	5.96-12.32
Canon TBA-FX8/TBA-120FR	CRP	mg/L	4.34	0.42	3.08-5.60
	Lp-PLA2	U/L	296.41	27.04	215.29-377.53
	MPO	ng/mL	54.77	5.12	39.41-70.13
	SAA	mg/L	8.83	0.80	6.43-11.23

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.