

Inflammation Markers Control – Level 2

Assignment of values (**REF** BYZKB0201 **LOT** 10H1104)

Instrument	Analyte	Unit	Inflammation Markers Control Level2		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	2.17	0.20	1.57-2.77
	IL-6	pg/mL	65.19	6.82	44.73-85.65
	CRP	mg/L	15.69	1.65	10.74-20.64
Roche cobas 6000 e 601	PCT	ng/mL	1.29	0.12	0.93-1.65
	IL-6	pg/mL	83.23	7.94	59.41-107.05
Beckman Coulter AU680	CRP	mg/L	17.42	1.90	11.72-23.12
	Lp-PLA2	U/L	408.64	40.37	287.53-529.75
	MPO	ng/mL	128.38	11.83	92.89-163.87
	SAA	mg/L	20.80	2.49	13.33-28.27
Canon TBA-FX8/TBA-120FR	CRP	mg/L	16.75	1.61	11.92-21.58
	Lp-PLA2	U/L	421.30	43.10	292.00-550.60
	MPO	ng/mL	116.81	10.48	85.37-148.25
	SAA	mg/L	19.96	1.99	13.99-25.93

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.