

Inflammation Markers Control – Level 3

Assignment of values (**REF** BYZKB0301 **LOT** 10H1104)

Instrument	Analyte	Unit	Inflammation Markers Control Level3		
			Target	SD	Range
AutoLumo A2000 Plus	PCT	ng/mL	24.83	2.46	17.45-32.21
	IL-6	pg/mL	491.01	44.43	357.72-624.30
	CRP	mg/L	30.65	2.89	21.98-39.32
Roche cobas 6000 e 601	PCT	ng/mL	17.44	1.77	12.13-22.75
	IL-6	pg/mL	592.29	61.02	409.23-775.35
Beckman Coulter AU680	CRP	mg/L	31.94	3.50	21.44-42.44
	Lp-PLA2	U/L	794.17	80.80	551.77-1036
	MPO	ng/mL	733.08	67.70	529.98-936.18
	SAA	mg/L	38.85	4.14	26.43-51.27
Canon TBA-FX8/TBA-120FR	CRP	mg/L	31.26	2.91	22.53-39.99
	Lp-PLA2	U/L	827.46	81.99	581.49-1073
	MPO	ng/mL	663.01	60.16	482.53-843.49
	SAA	mg/L	36.89	3.44	26.57-47.21

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.