

HPV-16/18 DNA Control– Level 1

Assignment of values (**REF** CFBB0101 **LOT** 10H1101)

Instrument	Assay kit	Analyte	Unit	HPV-16/18 DNA Control Level 1		
				Target	SD	Range
SLAN-96P	LIFERIVER	16-DNA	/	-	/	-
		18-DNA	/	-	/	-

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.