

HPV-16/18 DNA Control – Level 3

Assignment of values (**REF** CFBB0103 **LOT** 10H1101)

Instrument	Assay kit	Analyte	Unit	HPV-16/18 DNA Control Level 3		
				Target	SD	Range
SLAN-96P	LIFERIVER	16-DNA	/	12.08	/	10.08-14.08
		18-DNA	/	12.35	/	10.35-14.35

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.