

Molecular Respiratory Premium Control – Level 3

Assignment of values (**REF** CFBA0103 **LOT** 11H1101)

Instrument	Assay kit	Analyte	Unit	Molecular Respiratory Premium Control Level 3		
				Target	SD	Range
AutoMolec3000	Autobio	FluA	/	24.71	/	22.71-26.71
		FluB	/	23.56	/	21.56-25.56
		RSV	/	27.06	/	25.06-29.06
		HPIV-1	/	24.61	/	22.61-26.61
		HPIV-2	/	25.25	/	23.25-27.25
		HPIV-3	/	25.35	/	23.35-27.35
		MP	/	26.34	/	24.34-28.34
		CP	/	25.98	/	23.98-27.98
		ADV	/	25.87	/	23.87-27.87
		hMPV-A	/	26.49	/	24.49-28.49
		hMPV-B	/	25.67	/	23.67-27.67
SLAN-96P	ABT	FluA	/	26.92	/	24.92-28.92
		FluB	/	22.43	/	20.43-24.43
		RSV	/	26.19	/	24.19-28.19
		HPIV-1	/	23.67	/	21.67-25.67
		HPIV-3	/	23.11	/	21.11-25.11
		ADV	/	29.23	/	27.23-31.23
	BioGerm	RSV	/	26.41	/	24.41-28.41
		HPIV-1	/	24.11	/	22.11-26.11
		HPIV-2	/	27.28	/	25.28-29.28
		HPIV-3	/	25.82	/	23.82-27.82
		ADV	/	27.72	/	25.72-29.72
		hMPV-A	/	27.41	/	25.41-29.41
	SANSURE	FluA	/	25.57	/	23.57-27.57
		FluB	/	24.24	/	22.24-26.24
		RSV	/	26.40	/	24.40-28.40
		MP	/	19.70	/	17.70-21.70
		ADV	/	26.18	/	24.18-28.18
		RhV	/	23.48	/	21.48-25.48
	DAAN GENE	RhV	/	26.87	/	24.87-28.87
	LAND MEDICAL	RhV	/	24.75	/	22.75-26.75

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.