

Molecular Respiratory Premium Control Level 2

Assignment of values (**REF** CFBA0102 **LOT** 11H1101)

Instrument	Assay kit	Analyte	Unit	Molecular Respiratory Premium Control Level 2		
				Target	SD	Range
AutoMolec3000	Autobio	FluA	/	26.76	/	24.76-28.76
		FluB	/	25.74	/	23.74-27.74
		RSV	/	29.13	/	27.13-31.13
		HPIV-1	/	26.79	/	24.79-28.79
		HPIV-2	/	27.37	/	25.37-29.37
		HPIV-3	/	27.20	/	25.20-29.20
		MP	/	28.04	/	26.04-30.04
		CP	/	28.14	/	26.14-30.14
		ADV	/	27.86	/	25.86-29.86
		hMPV-A	/	28.56	/	26.56-30.56
		hMPV-B	/	27.79	/	25.79-29.79
SLAN-96P	ABT	FluA	/	28.93	/	26.93-30.93
		FluB	/	24.76	/	22.76-26.76
		RSV	/	28.32	/	26.32-30.32
		HPIV-1	/	26.04	/	24.04-28.04
		HPIV-3	/	24.84	/	22.84-26.84
		ADV	/	31.75	/	29.75-33.75
	BioGerm	RSV	/	28.36	/	26.36-30.36
		HPIV-1	/	26.80	/	24.80-28.80
		HPIV-2	/	29.41	/	27.41-31.41
		HPIV-3	/	27.58	/	25.58-29.58
		ADV	/	29.80	/	27.80-31.80
		hMPV-A	/	29.07	/	27.07-31.07
	SANSURE	FluA	/	27.70	/	25.70-29.70
		FluB	/	26.61	/	24.61-28.61
		RSV	/	28.45	/	26.45-30.45
		MP	/	22.26	/	20.26-24.26
		ADV	/	28.35	/	26.35-30.35
		RhV	/	25.65	/	23.65-27.65
	DAAN GENE	RhV	/	29.08	/	27.08-31.08
	LAND MEDICAL	RhV	/	26.86	/	24.86-28.86

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.