

Tumor Marker Speciality Control – Level 2

Assignment of values (**REF** CEA0402 **LOT** 17H1101)

Instrument	Analyte	Unit	Tumor Marker Control Level 2		
			Target	SD	Range
Auto Lumo A2000 plus	G-17	pmol/L	9.53	0.98	6.59 -12.47
	HER-2/neu	ng/mL	48.57	4.66	34.59 -62.55
	PIVKA-II	mAU/mL	443.23	24.45	364.38-511.08
	ProGRP	pg/mL	413.14	40.84	290.62 -535.66
	S100	ng/mL	3.23	0.33	2.24 -4.22
Auto Lumo A6200	G-17	pmol/L	9.70	0.73	7.51 -11.89
	HER-2/neu	ng/mL	49.00	4.50	35.5 -62.50
	PIVKA-II	mAU/mL	446.44	36.97	335.53 -557.35
	ProGRP	pg/mL	412.20	30.68	320.16 -504.24
	S100	ng/mL	3.24	0.22	2.58 -3.90
Auto Lumo A1860	G-17	pmol/L	9.51	0.81	7.08 -11.94
	HER-2/neu	ng/mL	49.55	4.87	34.94 -64.16
	PIVKA-II	mAU/mL	441.51	32.82	343.05 -539.97
	ProGRP	pg/mL	407.74	35.56	301.06 -514.42
	S100	ng/mL	3.05	0.22	2.39 -3.71

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.