

Tumor Marker Speciality Control – Level 1

Assignment of values (**REF** CEA0401 **LOT** 17H1101)

Instrument	Analyte	Unit	Tumor Marker Control Level 1		
			Target	SD	Range
Auto Lumo A2000 plus	G-17	pmol/L	3.58	0.36	2.50 -4.66
	HER-2/neu	ng/mL	12.94	1.12	9.58 -16.30
	PIVKA-II	mAU/mL	37.97	3.95	26.12 -49.82
	ProGRP	pg/mL	78.94	7.68	55.9 -101.98
	S100	ng/mL	0.37	0.03	0.28 -0.46
Auto Lumo A6200	G-17	pmol/L	3.60	0.29	2.73 -4.47
	HER-2/neu	ng/mL	11.78	1.02	8.72 -14.84
	PIVKA-II	mAU/mL	38.18	3.09	28.91 -47.45
	ProGRP	pg/mL	78.81	6.66	58.83 -98.79
	S100	ng/mL	0.39	0.03	0.30 -0.48
Auto Lumo A1860	G-17	pmol/L	3.71	0.31	2.78 -4.64
	HER-2/neu	ng/mL	12.02	0.92	9.26 -14.78
	PIVKA-II	mAU/mL	37.18	3.43	26.89 -47.47
	ProGRP	pg/mL	78.50	6.88	57.86 -99.14
	S100	ng/mL	0.39	0.03	0.30 -0.48

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.